



Locator and Locator Plus

**Cryogenic Storage Vessel with or without
Ultrasonic Level Monitor**

Operation Manual and Parts List LT509X9 (7018141) Rev. 6

Visit us online to register your warranty
www.thermoscientific.com/labwarranty

Thermo
SCIENTIFIC

Locator Catalog Number	Locator Model	Name
CY50925 w/o Monitor	8141	Locator Jr
CY50935 w/o Monitor	8143	Locator 4
CY50985 w/o Monitor	8197	Locator 6
CY50945 w/o Monitor	8145	Locator 8
CY509105 w/ Monitor	8201	Locator Jr.
CY509105CN w/ Monitor	8202	Locator Jr. China
CY509107 w/ Monitor	8246	Locator 4
CY509107CN w/ Monitor	8247	Locator 4 China
CY509113 w/ Monitor	8198	Locator 6
CY509110 w/ Monitor	8211	Locator 8
CY509110CN w/ Monitor	8212	Locator 8 China

Locator Plus Catalog Number	Locator Plus Model	Name
CY50925-70 w/o Monitor	8142	Locator Jr Plus
CY50935-70 w/o Monitor	8144	Locator 4 Plus
CY50985-70 w/o Monitor	8147	Locator 6 Plus
CY50945-70 w/o Monitor	8146	Locator 8 Plus
CY509106 w/ Monitor	8237	Locator Jr Plus
CY509106CN w/ Monitor	8238	Locator Jr Plus China
CY509108 w/ Monitor	8207	Locator 4 Plus
CY509108CN w/ Monitor	8208	Locator 4 Plus China
CY509109 w/ Monitor	8209	Locator 6 Plus
CY509109CN w/ Monitor	8210	Locator 6 Plus China
CY509111 w/ Monitor	8213	Locator 8 Plus
CY509111CN w/ Monitor	8214	Locator 8 Plus China

MANUAL NUMBER LT509X9 (7018141)

REV	ECR/ECN	DATE	DESCRIPTION	By
6	ECNM-000831	03/16/21	Removed Warranty caution from preface page ii, added Safety Precautions in Preface page iii, page 1-2, WEEE Compliance in page 9-1 and removed EUT statements in page 2-1. Updated Specifications for Environmental Conditions in page 2-1.	TFI
5	41864	11/07/17	Removed CE mark and info	bpq
4	40578	2/23/16	Clarified Do Not Roll warning	ccs
3	--	10/27/15	Updated symbols and labels, added cautions	ccs
2	31623	5/15/15	Added caution for moving vessel in Section 3	ccs
1	31528	9/15/14	Updated conformity specs and declaration	ccs
0	27564	1/31/14	Original - new PCBA design	ccs



Important Read this instruction manual. Failure to read, understand and follow instructions in this manual may result in damage to the unit, injury to operating personnel, and poor equipment performance. ▲




Caution All internal adjustments and maintenance must be performed by qualified service personnel. ▲



Warning Nitrogen gas can cause suffocation without warning. Store and use liquid nitrogen only in a well-ventilated place. As the liquid evaporates, the resulting gas displaces the normal air in the area. (The cloudy vapor that appears when liquid nitrogen is exposed to the air is condensed moisture, not the gas itself. The issuing gas is invisible.) In closed areas, excessive amounts of nitrogen gas reduces the concentration of oxygen and can result in asphyxiation. Because nitrogen gas is colorless, odorless, and tasteless, it cannot be detected by the human senses. Breathing an atmosphere that contains less than 19.5% oxygen can cause dizziness and quickly result in unconsciousness and death. Therefore, the use of oxygen monitoring equipment is strongly recommended. ▲



Warning Contact of liquid nitrogen with the skin or eyes may cause serious (freezing) injury. ▲



Warning Do not roll. Always keep liquid nitrogen vessel in an upright position. Exercise caution to prevent spilling and splashing. ▲

Material in this manual is for information purposes only. The contents and the product it describes are subject to change without notice. Thermo Fisher Scientific makes no representations or warranties with respect to this manual. In no event shall Thermo be held liable for any damages, direct or incidental, arising out of or related to the use of this manual.

Intended Use:

The Thermo Fisher Scientific cryogenic storage vessels are intended to store biological samples at cryogenic temperatures. These products are comprised of a vessel and a lid, with or without a level monitor.

Unintended Use:

These products are not intended for use within the patient environment.

When this manual is translated into other languages, the US English version is binding.

© 2021 Thermo Fisher Scientific. All rights reserved. Printed in U.S.A.



Important operating and/or maintenance instructions. Read the accompanying text carefully.



Potential electrical hazards. Only qualified persons should perform procedures associated with this symbol.



Extreme temperature hazards. Only qualified persons should perform procedures associated with this symbol.



Potential biological hazards. Proper protective equipment and procedures must be used.



This symbol when used alone indicates important operating instructions which reduce the risk of injury or poor performance of the unit.



CAUTION: This symbol, in the context of a CAUTION, indicates a potentially hazardous situation which if not avoided could result in minor to moderate injury or damage to the equipment. ▲



CAUTION: This indicates a situation which may result in property damage. ▲



Before installing, using or maintaining this product, please be sure to read the manual and product warning labels carefully. Failure to follow these instructions may cause the product to malfunction, which could result in injury or damage.



WARNING: Unauthorized repair of your freezer will invalidate your warranty. Contact Technical Service at 1-800-438-4851 for additional information. ▲

Do You Need Information or Assistance on Thermo Scientific Products?

If you do, please contact us 8:00 a.m. to 6:00 p.m. (Eastern Time) at:

1-740-373-4763

Direct

1-800-438-4851

Toll Free, U.S. and Canada

1-877-213-8051

FAX

<http://www.thermofisher.com>

Internet Worldwide Web Home Page

service.led.marietta@thermofisher.com

Tech Support Email Address

www.unitylabservices.com

Certified Service Web Page

Our **Sales Support** staff can provide information on pricing and give you quotations. We can take your order and provide delivery information on major equipment items or make arrangements to have your local sales representative contact you. Our products are listed on the Internet and we can be contacted through our Internet home page.

Our **Service Support** staff can supply technical information about proper setup, operation or troubleshooting of your equipment. We can fill your needs for spare or replacement parts or provide you with on-site service. We can also provide you with a quotation on our Extended Warranty for your Thermo Scientific products.

Whatever Thermo Scientific products you need or use, we will be happy to discuss your applications. If you are experiencing technical problems, working together, we will help you locate the problem and, chances are, correct it yourself...over the telephone without a service call.

When more extensive service is necessary, we will assist you with direct factory trained technicians or a qualified service organization for on-the-spot repair. If your service need is covered by the warranty, we will arrange for the unit to be repaired at our expense and to your satisfaction.

Regardless of your needs, our professional telephone technicians are available to assist you Monday through Friday from 8:00 a.m. to 6:00 p.m. Eastern Time. Please contact us by telephone or fax. If you wish to write, our mailing address is:



Thermo Fisher Scientific (Asheville) LLC
401 Millcreek Road, Box 649
Marietta, OH 45750

International customers, please contact your local Thermo Scientific distributor.

Table of Contents

Section 1	Safety Information	1-1
	Please Read	1-1
	General Usage	1-1
	Warnings	1-2
	Cautions	1-2
Section 2	General Specifications	2-1
	Environmental Conditions	2-1
Section 3	Unpacking	3-1
Section 4	Operation	4-1
	Filling Instructions	4-1
	Measuring Liquid Nitrogen Quantity	4-2
	Storing Samples in Vapor Phase	4-2
	General Cleaning Instructions	4-2
Section 5	Replacement Parts	5-1
Section 6	Liquid Nitrogen Level Monitor Operation	6-1
	Installation	6-1
	Remote Alarm	6-2
	Alarm Contact Specifications	6-2
	Microprocessor Functions	6-4
	Principles of Operation	6-4
	Level Monitor Operation	6-5
	Safety Feature	6-5
	Audible and Visual Alarms	6-5
Section 7	LN2 Level Monitor Troubleshooting and Parts List	7-1
	Ordering Procedures	7-2
Section 8	Warranty Information	8-1
Section 9	WEEE Compliance	9-1

Table of Contents

Section 1 Safety Information

This manual contains important operating and safety information. The user must carefully read and understand the contents of this manual prior to the use of this equipment.

Your Thermo Scientific cryogenic vessel has been designed with function, reliability, and safety in mind. It is the user's responsibility to install it in conformance with local electrical codes. For safe operation, please pay attention to the warnings and cautions throughout the manual.

Please Read

All cryopreservation vessels, regardless of who manufactures them, use a vacuum to provide the super insulative properties needed to keep nitrogen in a liquid form. They are very similar, in fact, to the Thermos® vessels you may have used to store coffee, soup or milk.

Remember how fragile they were? Since no vacuum is perfect or will last forever, we suggest that you monitor the consumption of liquid nitrogen used by your vessel on a regular basis, i.e. every 1-2 days. By monitoring, you may be able to anticipate subtle changes in consumption and possibly react to problems before they arise. You may opt to use a level monitor for this purpose (see page 6-1 for level monitor operation.).


Also, if your samples are super critical, one-of-a-kind, or irreplaceable, consider allocating your samples to 2 separate vessels. Remember the old adage about placing all your eggs into one basket....

General Usage

Do not use this product for anything other than its intended usage.

Use of the equipment in a manner not specified by the manufacturer may impair the protection provided by the equipment.

Warnings



- This unit is intended to be used by trained personnel in a well-ventilated environment.
- Liquid nitrogen is extremely cold; it boils at -196°C.
- Follow all necessary precautions and warnings relevant to applicable hazards such as frostbite, contact burn, asphyxiation, etc. in cryogenic area.
- To avoid injury due to frostbite or ruptured vials, use extreme care whenever handling liquid nitrogen, liquid nitrogen storage or transfer vessels, or any objects which have come in contact with liquid nitrogen.
 - Leave no areas of skin exposed.
 - Wear appropriate PPE where necessary; face shield, cryogenic gloves, cryogenic apron.
 - Do not tightly seal liquid nitrogen containers or prevent nitrogen gas from escaping.
 - Always handle ampules with tongs. They may explode when removed from the vessel.
 - Use extreme care to prevent spilling and splashing liquid nitrogen during transfer and removal of storage contents and holders.
 - Immediately remove any clothing or safety attire on which liquid nitrogen has been spilled or splashed.
 - Get immediate medical attention for any frostbite injuries due to liquid nitrogen.
- Use this product only in the way described in the product literature and in this manual. Before using it, verify that this product is suitable for its intended use. If the equipment is used in a manner not specified by the manufacturer, the protection provided by the equipment may be impaired.

Cautions



- Only use the power supply provided by the unit.
- Never overfill liquid nitrogen vessels. Liquid nitrogen level should never be above the tops of the racks inside the vessel. The excess weight may crack the neck and result in vacuum failure.
- **When inserting or removing racks, be careful not to come in contact with the neck tube area of the vessel. Remove or insert racks slowly in a vertical manner. Scratches on the neck tube area can cause premature vacuum failure.**
- **Do not tamper with or remove vacuum port** (covered by a black plastic cap on side of vessel); this will void warranty.
- Do not spill liquid nitrogen on vacuum port - this can cause vacuum failure.
- Always keep liquid nitrogen vessel in an upright position.
- Do not use this Locator vessel for transport.

Section 1
Safety Information

Section 2 General Specifications

	Locator 4	Locator 4 Plus	Locator 6	Locator 6 Plus	Locator 8	Locator 8 Plus	Locator JR	Locator JR Plus
Height (including lid) ¹	37.5" (95.2 cm)	39.5" (100.3 cm)	39.5" (100.3 cm)	39.5" (100.3 cm)	37.5" (95.2 cm)	39.5" (100.3 cm)	26.5" (67.3 cm)	28.9" (73.4 cm)
Diameter	22" (55.9 cm)	22" (55.9 cm)	26" (66 cm)	26" (66 cm)	22" (55.9 cm)	22" (55.9 cm)	22" (55.9 cm)	22" (55.9 cm)
LN2 Capacity, Liquid Phase Storage	111 L	121 L	184 L	184 L	111 L	121 L	60 L	71 L
Static LN2 Evaporation rate - L/Day ³	0.99	0.99	0.99	0.99	0.60	0.60	0.85	0.85
Liquid Phase Capacity (2mL Vial ² actually submerged in liquid nitrogen)	3600	4000	6000	6000	1800	2000	1600	2000

¹ The liquid nitrogen level monitor will add 1 inch to these heights. The transportation cart will add 4 inches to these heights.

² Using Nalgene System 100 cryoboxes with a 10 x 10 vial configuration (Nalgene System 100 vial or equivalent are needed).

³ Actual working time may vary due to current atmospheric conditions, container history, manufacturing tolerances, and any individual patterns of use.

Environmental Conditions

Operating temperature: -29°C to +60°C

10% to 95% relative humidity, non-condensing.

Altitude limit: 2,000 meters.

Storage: -29°C to +60°C; 10% to 95% relative humidity.

Intended for indoor use only.

Electrical (cover with monitors only)

Inputs 100-240 VDC, 0.08 - 0.04 amps, single phase

Frequency 47 - 63 Hz

Output 15 VDC (±10%), 0.2 amps, 3.0 watts

Section 2

General Specifications

Section 3 Unpacking

Before using your new Locator or Locator Plus vessel, carefully inspect the vessel prior to use. Check for signs of damage which may have occurred in shipment. It is advisable to fill (see filling instructions) all new units with liquid nitrogen and watch liquid nitrogen loss rate for a few days. If there are any problems, call Customer Service as soon as possible.

The warranty registration card enclosed with the vessel must be completed and returned to the factory within 30 days to make warranty effective. This information must include the serial number, which is located on the label on the vessel.

Note The most prevalent cause of failure of liquid nitrogen storage vessels is mechanical. The vessel necktube supports the full weight of the inner shell and all the liquid nitrogen it contains. A side blow to the vessel causes the inner shell to swing in a pendulum-like motion causing the necktube to be damaged. Any storage vessel which has been in an accident, has been dropped, or lowered to hit on its side, will tend to fail more rapidly than one that has not.

Caution Exercise caution when moving your Locator or Locator Plus vessel. Locator and Locator Plus cryobiological storage systems are not transportation vessels. Transport carts are designed for mobility within the lab, or lab to lab only. Moving full vessels long distances, over cracks in floor, thresholds, on inclined ramps or in elevators can cause premature vacuum failure. Avoid using excessive force when moving vessel; this can cause tipping, resulting in damage and/or spilling of liquid nitrogen. ▲



Note If samples must be transported under cryogenic conditions, consider Arctic Express line of Dry Shippers available from Thermo Scientific.

Section 3
Unpacking

Section 4 Operation



Caution Never overfill your Locator or Locator Plus vessel with liquid nitrogen. The liquid nitrogen level in your tank (with racks inserted) should never be above 20 inches (50 cm) for Locator 4 and Locator 8, 10 inches (25 cm) for Locator Jr., 22 inches (55.8 cm) for Locator 4, 6, and 8 Plus and 12 inches for Locator Jr. Plus. Filling the tank up to or above the bottom of the necktube may cause immediate or premature vacuum failure to occur. ▲



Caution When inserting or removing racks, be careful not to come in contact with the necktube area of the vessel. Remove or insert racks slowly in a vertical manner. Scratches on the necktube area can cause premature vacuum failure. ▲



Caution Do not spill liquid nitrogen on vacuum port (covered by a black plastic cap on side of vessel). This can cause vacuum failure. ▲

Filling Instructions

To avoid damage to your Locator or Locator Plus cryogenic storage vessel which may result in premature vacuum loss, it is important that the following procedure be used when adding liquid nitrogen to a warm vessel.

1. Add only a small amount of liquid nitrogen (5-10 liters) to a new or warm vessel.
2. Allow this small amount of liquid nitrogen to sit in the covered vessel for a minimum of 2 hours. This will limit stress caused by the sudden temperature change associated with adding liquid nitrogen to a warm vessel.
3. Add an additional 15 liters of liquid nitrogen to the vessel.
4. Allow vessel to sit for 48 hours and monitor liquid nitrogen consumption with a wooden yardsstick, optional dip measurer (part #180143) or Level Monitor.
5. Fill Locator or Locator Plus as indicated (see Caution in this section). Allow for displacement of liquid nitrogen when racks and boxes are inserted.
6. Insert and remove racks slowly. Allow liquid nitrogen to run out of boxes and off racks.

Measuring Liquid Nitrogen Quantity

1. Use a wooden yardstick or optional dip measurer (part #180143) to measure liquid nitrogen level. NEVER use a hollow tube or plastic dipstick to measure the liquid nitrogen level.
2. Level will be indicated by the frostline which develops when dipstick is removed.
3. Level Monitor can be used for constant measurements. See Section 6.

Storing Samples in Vapor Phase

1. Remove the 2 bottom-most cryoboxes from each rack.
2. Measure the height of 2 cryoboxes stacked on top of each other.
3. Fill vessel with liquid nitrogen to the height obtained from Step 2, using a yardstick to measure liquid level in vessel.
4. Place racks into vessel without the 2 bottom-most cryoboxes. Samples are now above the liquid nitrogen and are stored in the vapor phase.

Important A level monitor should always be used when you store your samples in the vapor phase because of the already low levels of liquid nitrogen. Refer to Section 5 for correct level monitor for your system. ▲

General Cleaning Instructions

Wipe exterior surfaces with lightly dampened cloth containing mild soap solution.

Section 5 Replacement Parts

Locator and Locator Plus Covers w/o Monitor		
Type	Standard Covers	Covers and Rack Kits (for units with lids made before 3/2010)
Locator Jr.	CV509X51 (321746)	1950514
Locator Jr. Plus	CV509X51 (321746)	1950515
Locator 4	CV509X51 (321746)	1950516
Locator 4 Plus	CV509X51 (321746)	1950517
Locator 6	CV509X50 (321651)	1950513
Locator 6 Plus	CV509X50 (321651)	1950513
Locator 8	CV509X23 (317227)	CV509X23 (317227)
Locator 8 Plus	CV509X23 (317227)	CV509X23 (317227)

Section 5

Replacement Parts

Catalog #	Rack Description
HR509X15A (317050)	For Locator Jr (8141, 8201, 8202) with 4 - 2 inch boxes
HR509X2A (317161)	For Locator 4 (8143, 8246, 8247) with 9 - 2 inch boxes
HR509X19A (315903)	For Locator Jr (8141, 8201, 8202) with 2 - 4 inch boxes
HR509X20A (315905)	For Locator 4 (8143, 8246, 8247) with 5 - 4 inch boxes
HR509X15A-70 (317148)	For Locator Jr Plus (8142, 8237, 8238) with 5 - 2 inch boxes
HR509X2A-70 (317175)	For Locator 4 Plus (8144, 8207, 8208) with 10 - 2 inch boxes
HR509X31A-70 (317192)	For Locator 6 and 6 Plus (8147, 8197, 8198, 8209, 8210) with 10 - 2 inch boxes
HR509X1A (317205)	For Locator 8 (8211, 8212, 8145) with 9 - 2 inch boxes (25 cell)
HR509X1A-70 (317217)	For Locator 8 Plus (8146, 8213, 8214) with 10 - 2 inch boxes (25 cell)
HR509X19A-70 (315904)	For Locator Jr. Plus (8142, 8237, 8238) with 2 - 4 inch boxes + 1 - 2 inch box
HR509X20A-70 (315906)	For Locator 4 Plus (8144, 8207, 8208) with 5 - 4 inch boxes + 1 - 2 inch box
HR509X32A-70 (315907)	For Locator 6 and 6 Plus (8147, 8197, 8198, 8209, 8210) with 5 - 4 inch boxes + 1 - 2 inch box

Section 6 Liquid Nitrogen Level Monitor Operation

The liquid nitrogen (LN₂) level monitor provides a constant indication of the LN₂ level and notifies you to low level conditions through audible and visual alarms. It alerts you when your vessel needs filling. The level monitor is shipped mounted directly to the appropriate Locator or Locator Plus cover. It can be wired to a remote alarm system to alert you to problems when you are not in your laboratory. This equipment was intended to be operated with an external power supply provided by the manufacturer. It is designed for use with Thermo Scientific LN₂ vessels only.



Caution Use this product for its intended usage only. ▲



Warning Potential electrical hazards. Only qualified persons should perform procedures associated with this symbol. ▲

Installation

Note If the level monitor was purchased as part of a Locator or Locator Plus system, skip to Step 3. If it is a retrofit for an existing Locator or Locator Plus, begin with Step 1. ▲

1. Remove the level monitor and Locator or Locator Plus cover from box.
2. Remove present cover from your Locator or Locator Plus and replace with your new level monitor and cover.
3. Check the electrical specifications on the power supply for electrical data and plug in to appropriately grounded receptacle.

Level Monitor Remote Alarm

1. Disconnect the liquid nitrogen level monitor from the power supply. Disconnect power to your alarm system.
2. Remove the three screws securing the liquid nitrogen level monitor's cover. Pull the cover up and out of the way (see Figure 6-1).

Connection to Remote Alarms

When connected to your remote alarm system, the LN₂ level monitor will alert you to a problem even when your laboratory is unoccupied. The unit's alarm contacts are activated at the same time the audible alarm sounds (see Microprocessor Functions).

3. Remove the black plug from the side of the cover. Insert the wires for the remote alarm through this hole.
4. Connect the remote alarm wires as shown in Figures 6-1, 2 and 3.
5. Replace the LN₂ level monitor's cover. Reconnect power to the LN₂ level monitor and to your remote alarm system.

Note The LN₂ level monitor can be configured with either a normally open contact or a normally closed contact. Which mode is currently operative depends on the placement of a single jumper on the circuit board. For a normally open contact, the jumper should be placed on J6. For a normally closed contact, the jumper should be placed on J4 (see Figure 6-2). Verify that the jumper is set appropriately for your remote alarm system. ▲

Alarm Contact Specifications

The LN₂ level monitor's alarm contact is a Photomos type relay.

Contact Form: Type 1 form B

Load Voltage Maximum: 24 VDC Max

Load Current Maximum: 120 mA

Typical On Resistance: Maximum 26 Ω




Figure 6-1. Connect to Remote Alarm




Figure 6-2. Jumper Location

Principles of Operation

The innovative Thermo LN₂ Level Monitor utilises ultrasonic sound waves to sense the level of liquid nitrogen inside a Locator or Locator Plus cryogenic storage vessel. A microprocessor control provides continuous indication of the liquid nitrogen level inside the Locator or Locator Plus, with accuracy to $\pm 2''$. Audible and visual alarms also provide an early warning of low level conditions.

Ultrasonic sound waves sense the level of liquid nitrogen inside the Locator or Locator Plus vessel.

By using ultrasonic sound waves to sense the level of liquid nitrogen, no physical probe is required to sense the level of liquid nitrogen. Conventional devices use probes which must be in physical contact with the liquid. This physical contact increases liquid nitrogen consumption. There is NO increase in liquid nitrogen consumption by using the Thermo level monitor.

Micropocessor Functions

The LED display is graduated in 8 segments and provides quick and reliable visual indication of the level of liquid nitrogen inside the Locator or Locator Plus vessel, from Full to Empty, similar to a fuel gauge on an automobile.




Figure 6-3. Remote Alarm Connections Schematic

Audible and Visual Alarms

1. “LOW LEVEL” indicator light illuminates continuously when liquid nitrogen level is below the approximate 2" level. In this situation, the level of liquid nitrogen is low, and you are alerted to refill the vessel as soon as possible.
2. An AUDIBLE and remote alarm activates continuously when liquid nitrogen is below the approximate 1" level inside the Locator or Locator Plus vessel, additionally, “LOW LEVEL” light will continue to flash. In this situation, the liquid nitrogen is dangerously low, and you are alerted to refill the vessel immediately.
3. The AUDIBLE and remote alarms activate continuously, the 8-segment LED display and “Low LEVEL” lights flash, to indicate an emergency event. This alarm event occurs when the Locator or Locator Plus vessel is empty of liquid nitrogen, or the monitor has an error and is incapable of detecting LN₂ levels.

Safety Feature

The microprocessor must be reset each time the cover is removed from the Locator or Locator Plus. **Before removing the cover, the RESET button must be pressed**, illuminating the RESET light. While in this Reset/Hold mode, the monitor is not sensing. After five minutes in the Reset/Hold mode, an audible and remote alarm will activate, alerting you to deactivate the Reset/Hold mode when the cover is replaced on the vessel.

Level Monitor Operation

Once the cover with the monitor has been placed on the vessel and plugged in, wait about 15 minutes for the system to equilibrate before pressing the Reset button for an accurate measurement of the liquid nitrogen to be indicated. This equilibration process will occur each time the cover is removed. Each time a vessel is opened or moved, the stabilized chamber temperature and/or liquid level is disrupted. The system requires a stable temperature and liquid level to determine an accurate liquid nitrogen level.

If you have just purchased a new Locator or Locator Plus along with your monitor, allow at least 1 day for the liquid nitrogen to equilibrate the vessel before plugging in the unit. This will avoid false alarms.

Note The Level Monitor may take up to 10 minutes to establish a reading after the unit is placed on the cryovessel, or the cryovessel is moved. Remember to press “Reset” after placing unit on the vessel. The green LED on the switch should be “OFF.” ▲

Section 6

LN2 Level Monitor Operation

Section 7 LN₂ Level Monitor Troubleshooting and Parts List



Caution - Preventative Maintenance Depending on environmental conditions and the number of lid openings, the lid should be periodically inspected for ice and frost buildup. Refer to chart below. ▲

Problem	Cause	Solution
Level monitor alarm sounds intermittently. Unit measuring liquid level incorrectly.	Ice or frost buildup on the transducer.	Remove the cover w/ monitor from vessel and place in room for approximately two hours, or until no ice is visible on the transducer or tube. NOTE: Do not use an external heat source (i.e. heat gun) to accelerate the thawing process as it can damage the cork. Make sure lid is standing upright on the cork to allow moisture to drain out of cover w/ monitor. Wipe any moisture away from the inside of the tube and face of the transducer with a clean soft cloth. Place cover w/ monitor back on unit. (Remember to press "Reset" button.)
Level monitor measuring "Full" after removing samples and replacing lid.	Fog in top part (neck area) of cryogenic vessel.	The density of the fog that is formed after removing the lid and accessing samples is dependent on the humidity of the room where the cryovessel is located. The fog dissipates after about fifteen minutes and at this time, the level monitor will take an accurate reading of the liquid nitrogen level.

Ordering Procedures

Refer to the Specification Plate for the complete model number, serial number, and series number when requesting service, replacement parts or in any correspondence concerning this unit.

All parts listed herein may be ordered from the Thermo Scientific dealer from whom you purchased this unit or can be obtained promptly from the factory. When service or replacement parts are needed, check first with your dealer. If the dealer cannot process your request, then contact Technical Services.

Prior to returning any materials, contact our Technical Services Department for a "Return Materials Authorization" number (RMA). Material returned without an RMA number will be refused.

Table 7-1. Parts List By Model

PART	LOCATOR JR	LOCATOR JR PLUS	LOCATOR 4	LOCATOR 4 PLUS	LOCATOR 6	LOCATOR 6 PLUS	LOCATOR 8	LOCATOR 8 PLUS
Cover with Monitor and Power Supply *(316306)	CN509X16 (316308)	CN509X17 (316374)	CN509X17-70 (316376)	CN509X15 (316304)	CN509X15 (316304)	CN509X15 (316304)	CN509X7 (316137)	CN509X7-70 (316138)
Cover with Monitor and Power Supply (China) **(316307)	CN509X16CN (316309)	CN509X17CN (316375)	CN509X18CN (316375)	CN509X19CN (316377)	CN509X15CN (316305)	CN509X15CN (316305)	CN509X7CN (316139)	CN509X20CN (316390)
Cover and Rack Kits* (for units with lids made before 03/2010)	1950503	1950505	1950507	1950509	1950501	1950501	CN509X7 (316137)	CN509X7-70 (316138)
Cover and Rack Kits (China) ** for units with lids made before 03/2010	1950504	1950506	1950508	1950510	1950502	1950502	CN509X7CN (316139)	CN509X20CN (316390)
Cover	CV509X13	CV509X13	CV509X13	CV509X13	CV509X13	CV509X13	CV509X13	CV509X13
Power Supply w/ connectors* 90-240VAC	TN509X2 (317133)	* includes US, EU, UK and AUS connectors						
Power Supply** 90-264VAC (China)	TNX136 (317143)							
Vent Port Cap	1950343							