

Thermo Scientific NanoDrop 8000 Spectrophotometer

NanoDrop 8000

| | |
|------------------------------|--|
| Instrument Type: | Spectrophotometer |
| Minimum Sample Size: | 1 μ l |
| Sample Number: | up to 8 |
| Path Length: | 1mm (auto-ranging to 0.2 mm) |
| Light Source: | Xenon flash lamp |
| Detector Type: | 2048-element linear silicon CCD array |
| Wavelength Range: | 220 – 750 nm |
| Wavelength Accuracy: | 1 nm |
| Spectral Resolution: | 3 nm (FWHM at Hg 546 nm) |
| Absorbance Precision: | 0.003 (1 mm path) |
| Absorbance Accuracy: | 2% (at 0.76 at 257 nm) |
| Absorbance Range: | 0.02 – 75 (10 mm equivalent) |
| Detection Limit: | 2.5 ng/ μ l (dsDNA) |
| Maximum Concentration: | 3,700 ng/ μ l (dsDNA) |
| Measurement Time: | < 20 seconds |
| Footprint: | 24 x 32 cm |
| Weight: | 3.4 kg |
| Sample Pedestal Material | |
| of Construction: | 303 stainless steel and quartz fiber |
| Operating Voltage: | 12 vdc |
| Operating Power Consumption: | 30 W |
| Standby Power Consumption: | 6 W |
| Software Compatibility: | Windows [®] 2000 XP and Vista [™] (32 bit) |

All NanoDrop instruments are approved to CE and UL/CSA standards.

NanoDrop Products Patented Retention System

All NanoDrop products utilize a unique technology that allows a sample to be pipetted directly onto an optical measurement surface. The system uses inherent surface tension to hold a micro-volume sample in place during the measurement cycle. Once the measurement is complete, the surfaces are simply wiped with a lint-free lab wipe.



Our trial program allows you to try an instrument in your lab with your own samples—completely free of charge. Visit www.nanodrop.com to request your free trial instrument.*

* Available only in US and Canada

© 2009 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific Inc. and its subsidiaries.

Thermo Fisher Scientific
NanoDrop Products
3411 Silverside Road, Bancroft Building
Wilmington, DE 19810 USA

www.nanodrop.com

302-479-7707

Thermo
SCIENTIFIC